

2	<u>Chair for Hemodialysis Machine</u>
---	---------------------------------------

No.	Item Specifications	Compliance	Fill Your Specifications
*	Manufacturer Name:		
**	Country of Origin:		
***	Model No.:		
1	The chair must be designed to hemodialysis and conform to international specifications and standards		
2	To work by electrical current system and be identical to the current system in Sudan (220 or 240) volt (50 or 60) Hz		
3	Operating system with remote control		
4	The possibility of operating with a dry battery (back up battery)		
5	It contains number of different movements for (leg / head / back)		
6	To bear the weight of more than (200) kg		
7	The (length width height) of chair must be taken into account(more than) (1900—600--900)mm		
8	The materials from which the chair frame is made must be solid unbreakable and easy to clean		
9	The materials used for upholstery and dressing must be flexible , comfortable and easy to clean medical materials		