

AMENDMET No. 1 TO NMSF TENDER NO. 13/2022
“Technical Specifications”

1	Section:	High speed oxygen cylinder filling system with filling ramp for six cylinders:
	<u>Amendment</u>	<p>The system should be with high speed high pressure oil free compressor. The compressor compresses & fills the low pressure oxygen generated from oxygen generator into different size of cylinders at high pressure 150 bar.</p> <p>* The filling capacity for 2 cylinders of 6m³/h.</p> <p>Low power consumption & energy saving. Suction oxygen pressure: 6-20 PSI Oxygen Product Flow rate: 5-10 LPM (The Minimum product flow should be at least 15 LPM & The maximum capacity of booster should not exceed 20 LPM) Filling ramp for 6 cylinders</p>