



## **National Medical Supplies Fund**

## **CLARIFICATIONS No. 1 To NMSF TENDER 01-2016**

	QUESTION:	On the tender document, there is "Auto-Positioning is required". I just want to
	QCLSTIOI.	clarify with this. What is the meaning of "Auto-Positioning"? I need you to
1		explain this part clearly if you do not mind.
	NMSF Replay:	It is very clear in specification
	QUESTION	We could not find specification of item below:
		1. 20 3200017 Operating Table
2		2. 35 3030042 Oxygen cylinder Regulator
		3. 36 3030044 N2O cylinder Regulator
	NMSF Replay:	1- See Attached file.
		2- & 3- Suitable for Anesthesia Machines.
	QUESTION:	We noticed that quantities sheet put need of # 500 hematology analyzers, but
3		when we took a look into specs sheet there were two separate specifications for
		Hematology analyzer (one for 3 differential and the other for 5 differential!!).
		do you meant your requested quantity are 500 each? or 250 each? or something
		else.
	NMSF Replay:	We need Only 3 part differential Hematology analyzers.
4	QUESTION:	I'd like to inquire about two items in operation room & ICU Department :
		1. item No 19 - Emergency Trolley in items list Vs patient trolley in
		specification list
		2. item No 20 - Operating Table specification missing in specification list.
	NMSF Replay:	See Attached files.
	QUESTION:	I need ask about:
		Track costomy, Tuke motel
		Tracheostomy Tube metal.
5		Tracheostomy Tube PVC with Cuff.
		please send me the Specification.
	NMSF Replay:	1- Jackson silver coating.
		2- Standard.
	QUESTION:	We would like to pay your attention that so many items specification not clear
		and may be misunderstanding
		1- Oxynn Mask Adult paedatric and neonate. How come in 3 ply.
		2- Endotracheal tubes. You mean PVC material. Or. silicon. ???
		3- Lab coats. Material. 100%cotton. Or. Mixed polyester with cotton and
_		what is need.
6		4- spinal needle. What type Quinck. Or pencil point tip.
		5- surgical gowns and patient gowns. What material need. SMS or non
		woven. And what weight need gram 35gram or 40.
	NIMOE D. 1	6- surgical stapler what size and what type.
	NMSF Replay:	1- 3-ply: Means special type of material.
		2- Endotracheal tubes we mean Silicon Type as in specificatioc.
		3- Lab Coats: we need minimum 60% Cotton and maximum 40%

		Polyester.
		4- Spinal Needle: We need Quincke type.
		5- Non woven – 40 gram.
7	QUESTION:	please we have some question about ultra sound specification item: Echo Ultrasound code: 3100041 question(1): point 3.1:  3-11 MHz Electronics Phased Array Probe for Vascular applications- 01 it shoud be Linear probe?? for vascular application!!  question(2): point 4.5: Harmonic Angio. we need to know if this is contrast Harmonic? question(3): point 4.27: Facility of Real time perfusion studies We need to check if is contrast imaging for cardiac?
	NMSF Replay:	As Tender Specifications
8	QUESTION:	Checking the specifications of strilization, but I have some questions.  - Autoclave (50L): it's not clear what exactly needed for (autoclaves are required for sterilizing an object in high temperature and high pressure steam): please inform for what it will be used!  - Autoclave (18L): same as above.
	NMSF Replay:	As Tender Specifications
	QUESTION:	السادة الافاضل نرجو من سيادتكم توضيح الاتي: (1) كمية الاجهزة المطلوبة من جهاز فحص التجلط-Coagulometer (Semi-automated, Fully  Automated)
9		2) ضرورة التقيد بالمواصفات الفنية المطلوبة نسبة لوجود بعض الفروقات.
	NMSF Replay:	1- We need only Semi-automated Coagulometer
		أى صنف يتحصل على 80% من المواصفة الفنية يدخل في المنافسة.   - 2
10	QUESTION:	Regarding the Glucometer listed in general unified tender for medical devices number 1/2016 please specify the requested glucometer detect which kind of enzymes GDD or GDH
	NMSF Replay:	As Tender Specifications
11	QUESTION:	Referring to your tender # 01/2016, Would you please clarify the size required (how many blood bags) of item # 26 on Laboratory Section as per your items list as bellow:  26 4060041 Platelet agitator and Incubator
	NMSF Replay:	Minimum 48 platelet bags

	QUESTION:	I would like to note that I couldn't find the specifications and required
	Q0201101	documentation for the below items for the Medical Equipments Tender
		(2016/1). Please provide me with them.
		1.Operating table
		2.DHS set
		3.Austin Moore set
		4.Sterilization Drum (3 Pcs)
12		5.Dressing Tray
		6.Ear syringe
		7.Sims Retractor
		8.Artery Forceps
		9.Kidney Dish – Large
		10. Kale pot (Bowl)
	NMSF Replay:	1- See Attached files.
		The rest is Standard Specification.